

## DECLARATION OF CONFORMITY

Product Identification : PM500A - Analog Input Graphical Panel Meters  
PM500D - Digital Input Graphical Panel Meters

Models : PM500A0301600F00, PM500A0400800F00, PM500D0301600F00  
PM500D0400800F00, PMA000CP00800000.

Option Modules:  
PMM000CM23200000, PMM000CM200000, PMM000I0AN000000  
PMM000I0RL200000, PMM000I0RL400000, PMM000PWACP00000.

Manufacturer : Red Lion Controls 20 Willow Springs Circle York, PA 17406

EU Representative : Red Lion Controls BV  
Spoetnik 20 Phone: +31 (0) 33-4723-225  
NL - 3824 MG Amersfoort Fax: +31 (0) 33-4893-793  
Netherlands [europa@redlion.net](mailto:europa@redlion.net)

UK Representative : Servomex Group Ltd.  
Jarvis Brook, Phone: +44 (0) 1892 652 181  
Crowborough,  
East Sussex, TN6 3FB, UK [customer.service@redlion.net](mailto:customer.service@redlion.net)

Function : PM500A Analog Input Graphical Panel Meters  
PM500D Digital Input Graphical Panel Meters and Option Modules

EMC Standards Used : EN 61326-1: 2013 - Electrical Equipment for Measurement, Control and Laboratory use;  
Immunity to Industrial Locations, Emissions to Class A

Safety Standards Used : Safety requirements for electrical equipment for measurement control and laboratory use  
EN 61010-1: 2010 - General Requirements  
EN 61010-2-030: 2010 - Particular requirements for testing and measuring circuits  
EN 61010-2-201: 2013 - Particular requirements for control equipment


### Means of CE Conformity

EMC Directive 2014/30/EU, LVD Directive 2014/35/EU, RED Directive 2014/53/EU and RoHS Directive 2011/65/EU

### Means of UK Legislation Conformity

Electromagnetic Compatibility Regulations 2016 (S.I. 2016/1091), Electrical Equipment (Safety) Regulations 2016 (S.I. 2016/1101),  
Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (S.I. 2012/3032)

DOC# : 2001C  
Issue Date : January 10, 2022



President (Jack Lee)